

## OIV-MA-AS2-12 Method for $^{18}\text{O}/^{16}\text{O}$ isotope ratio determination of water in wines and must

Type II method

### 1. Scope

The method describes the determination of the  $^{18}\text{O}/^{16}\text{O}$  isotope ratio of water from wine and must after equilibration with  $\text{CO}_2$ , using the isotope ratio mass spectrometry (IRMS).

### 2. Reference standards

ISO 5725:1994: Accuracy (trueness and precision) of measurement methods and results: Basic method for the determination of repeatability and reproducibility of a standard measurement method.

- V-SMOW: Vienna-Standard Mean Ocean Water ( $^{18}\text{O}/^{16}\text{O} = R_{V-SMOW} = 0.0020052$ )
- GISP Greenland Ice Sheet Precipitation
- SLAP Standard Light Antarctic Precipitation

### 3. Definitions

$^{18}\text{O}/^{16}\text{O}$  Isotope ratio of oxygen 18 to oxygen 16 for a given sample

$\delta^{18}\text{O}_{V-SMOW}$  Relative scale for the expression of the isotope ratio of oxygen 18 to oxygen 16 for a given sample.  $\delta^{18}\text{O}_{V-SMOW}$  is calculated using the following equation:

$$\delta^{18}\text{O}_{V-SMOW} = \left[ \frac{\left( \frac{^{18}\text{O}}{^{16}\text{O}} \right)_{\text{sample}} - \left( \frac{^{18}\text{O}}{^{16}\text{O}} \right)_{\text{standard}}}{\left( \frac{^{18}\text{O}}{^{16}\text{O}} \right)_{\text{standard}}} \right] \times 1000 [‰]$$

using the V-SMOW as standard and as reference point for the relative  $\delta$  scale.

- BCR: Community Bureau of Reference
- IAEA: International Atomic Energy Agency (Vienna, Austria)
- IRMM: Institute for Reference Materials and Measurements

# COMPENDIUM OF INTERNATIONAL METHODS OF WINE AND MUST ANALYSIS


## Method for $^{18}\text{O}/^{16}\text{O}$ isotope ratio determination of water in wines and must (Type-II)

---

- IRMS: Isotope Ratio Mass Spectrometry
- m/z: mass to charge ratio
- NIST: National Institute of Standards & Technology
- RM: Reference Material

### 4. Principle

The technique described thereafter is based on the isotopic equilibration of water in samples of wine or must with a  $\text{CO}_2$  standard gas according to the following isotopic exchange reaction:



After equilibration the carbon dioxide in the gaseous phase is used for analysis by means of Isotopic Ratio Mass Spectrometry (IRMS) where the  $^{18}\text{O}/^{16}\text{O}$  isotopic ratio is determined on the  $\text{CO}_2$  resulting from the equilibration.

### 5. Reagents and materials

The materials and consumables depend on the method used (see chapter 6). The systems generally used are based on the equilibration of water in wine or must with  $\text{CO}_2$ .

The following reference materials, working standards and consumables can be used:

#### 5.1. Reference materials

| Name            | issued by   | $\delta^{18}\text{O}$ versus V-SMOW |
|-----------------|-------------|-------------------------------------|
| V-SMOW, RM 8535 | IAEA / NIST | 0 ‰                                 |
| BCR-659         | IRMM        | -7.18 ‰                             |
| GISP, RM 8536   | IAEA / NIST | -24.78 ‰                            |
| SLAP, RM 8537   | IAEA / NIST | -55.5 ‰                             |

#### 5.2. Working Standards

5.2.1. Carbon dioxide as a secondary reference gas for measurement (CAS 00124-38-9).

5.2.2. Carbon dioxide used for equilibration (depending on the instrument this gas could be the same as 5.2.1 or in the case of continuous flow systems cylinders

# COMPENDIUM OF INTERNATIONAL METHODS OF WINE AND MUST ANALYSIS

## Method for $^{18}\text{O}/^{16}\text{O}$ isotope ratio determination of water in wines and must (Type-II)

---

containing gas mixture helium-carbon dioxide can also be used)

Working Standards with calibrated  $\delta^{18}\text{O}_{V-SMOW}$  values traceable to international reference materials.

### 5.3. Consumables

Helium for analysis (CAS 07440-59-7)

## 6. Apparatus

### 6.1. Isotope ratio mass spectrometry (IRMS)

The Isotope ratio mass spectrometer (IRMS) enables the determination of the relative contents of  $^{18}\text{O}$  of  $\text{CO}_2$  gas naturally occurring with an internal accuracy of 0.05%. Internal accuracy here is defined as the difference between 2 measurements of the same sample of  $\text{CO}_2$ .

The mass spectrometer used for the determination of the isotopic composition of  $\text{CO}_2$  gas is generally equipped with a triple collector to simultaneously measure the following ion currents:

- $m/z = 44$  ( $^{12}\text{C}^{16}\text{O}^{16}\text{O}$ )
- $m/z = 45$  ( $^{13}\text{C}^{16}\text{O}^{16}\text{O}$  and  $^{12}\text{C}^{17}\text{O}^{16}\text{O}$ )
- $m/z = 46$  ( $^{12}\text{C}^{16}\text{O}^{18}\text{O}$ ,  $^{12}\text{C}^{17}\text{O}^{17}\text{O}$  and  $^{13}\text{C}^{17}\text{O}^{16}\text{O}$ )

By measuring the corresponding intensities, the  $^{18}\text{O}/^{16}\text{O}$  isotopic ratio is determined from the ratio of intensities of  $m/z = 46$  and  $m/z = 44$  after corrections for isobaric species ( $^{12}\text{C}^{17}\text{O}^{17}\text{O}$  and  $^{13}\text{C}^{17}\text{O}^{16}\text{O}$ ) whose contributions can be calculated from the actual intensity observed for  $m/z = 45$  and the usual isotopic abundances for  $^{13}\text{C}$  and  $^{17}\text{O}$  in Nature.

The isotope ratio mass spectrometry must either be equipped with:

- a double introduction system (dual inlet system) to alternately measure the unknown sample and a reference standard.
- or a continuous flow system that transfers quantitatively the  $\text{CO}_2$  from the sample vials after equilibration but also the  $\text{CO}_2$  standard gas into the mass spectrometer.

### 6.2. Equipment and Materials

All equipments and materials used must meet stated requirements of the used method

# COMPENDIUM OF INTERNATIONAL METHODS OF WINE AND MUST ANALYSIS

## Method for $^{18}\text{O}/^{16}\text{O}$ isotope ratio determination of water in wines and must (Type-II)

---

/ apparatus (as specified by the manufacturer). However, all equipments and materials can be replaced by items with similar performance.

6.2.1. Vials with septa appropriate for the used system

6.2.2. Volumetric pipettes with appropriate tips

6.2.3. Temperature controlled system to carry out the equilibration at constant temperature, typically within  $\pm 1\text{ }^\circ\text{C}$

6.2.4. Vacuum pump (if needed for the used system)

6.2.5. Autosampler (if needed for the used system)

6.2.6. Syringes for sampling (if needed for the used system)

6.2.7. GC Column to separate  $\text{CO}_2$  from other elementary gases (if needed for the used system)

6.2.8. Water removal device (e.g. cryo-trap, selective permeable membranes)

### 7. Sampling

Wine and must samples as well as reference materials are used for analysis without any pre-treatment. In the case of the possible fermentation of the sample, benzoic acid (or another anti-fermentation product) should be added or filtered with a with a  $0,22\text{ }\mu\text{m}$  pore diameter filter.

Preferably, the reference materials used for calibration and drift-correction should be placed at the beginning and at the end of the sequence and inserted after every ten samples.

### 8. Procedure

The descriptions that follow refer to procedures generally used for the determination of the  $^{18}\text{O}/^{16}\text{O}$  isotopic ratios by means of equilibration of water with a  $\text{CO}_2$  working standard and the subsequent measurement by IRMS. These procedures can be altered according to changes of equipment and instrumentation provided by the manufacturers as various kind of equilibration devices are available, implying various conditions of operation. Two main technical procedures can be used for introduction of  $\text{CO}_2$  into the IRMS either through a dual inlet system or using a continuous flow system. The description of all these technical systems and of the corresponding conditions of operation is not possible. Note: all values given for volumes, temperatures, pressures and time periods are only indicative. Appropriate values must be obtained from specifications provided by the manufacturer and/or determined experimentally.

8.1. Manual equilibration

A defined volume of the sample/standard is transferred into a flask using a pipette. The flask is then attached tightly to the manifold.

Each manifold is cooled down to below  $-80\text{ }^{\circ}\text{C}$  to deep-freeze the samples (manifold equipped with capillary opening tubes do not require this freezing step). Subsequently, the whole system is evacuated. After reaching a stable vacuum the gaseous  $\text{CO}_2$  working standard is allowed to expand into the various flasks. For the equilibration process each manifold is placed in a temperature controlled water-bath typically at  $25^{\circ}\text{C}$  ( $\pm 1\text{ }^{\circ}\text{C}$ ) for 12 hours (overnight). It is crucial that the temperature of the water-bath is kept constant and homogeneous.

After the equilibration process is completed, the resulting  $\text{CO}_2$  is transferred from the flasks to the sample side bellow of the dual inlet system. The measurements are performed by comparing several times the ratios of the  $\text{CO}_2$  contained in the sample side and the standard side ( $\text{CO}_2$  reference standard gas) of the dual inlet. This approach is repeated till the last sample of the sequence has been measured.

### 8.2. Use of an automatic equilibration apparatus

A defined volume of the sample/standard is transferred into a vial using a pipette. The sample vials are attached to the equilibration system and cooled down to below  $-80\text{ }^{\circ}\text{C}$  to deep-freeze the samples (systems equipped with capillary opening tubes do not require this freezing step). Subsequently, the whole system is evacuated.

After reaching a stable vacuum the gaseous  $\text{CO}_2$  working standard is expanded into the vials. Equilibrium is reached at a temperature of typically  $22 \pm 1\text{ }^{\circ}\text{C}$  after a minimum period of 5 hours and with moderate agitation (if available). Since the equilibration duration depends on various parameters (e.g. the vial geometry, temperature, applied agitation ...), the minimum equilibrium time should be determined experimentally.

After the equilibration process is completed, the resulting  $\text{CO}_2$  is transferred from the vials to the sample side bellow of the dual inlet system. The measurements are performed by comparing several times the ratios of the  $\text{CO}_2$  contained in the sample side and the standard side ( $\text{CO}_2$  reference standard gas) of the dual inlet. This approach is repeated till the last sample of the sequence has been measured.

### 8.3. Manual preparation manual and automatic equilibration and analysis with a dual inlet IRMS

A defined volume of sample / standard (eg.  $200\text{ }\mu\text{L}$ ) is introduced into a vial using a pipette. The open vials are then placed in a closed chamber filled with the  $\text{CO}_2$  used for equilibration (5.2.2). After several purges to eliminate any trace of air, the vials are closed and then placed on the thermostated plate of the sample changer. The equilibration is reached after at least 8 hours at  $40\text{ }^{\circ}\text{C}$ . Once the process of

equilibration completed, the  $\text{CO}_2$  obtained is dried and then transferred into the sample side of the dual inlet introduction system. The measurements are performed by comparing several times the ratios of the  $\text{CO}_2$  contained in the sample side and the standard side ( $\text{CO}_2$  reference standard gas) of the dual inlet. This approach is repeated till the last sample of the sequence has been measured.

8.4. Use of an automatic equilibration apparatus coupled to a continuous flow system  
A defined volume of the sample/standard is transferred into a vial using a pipette. The sample vials are placed into a temperature controlled tray.

Using a gas syringe the vials are flushed with mixture of He and  $\text{CO}_2$ . The  $\text{CO}_2$  remains in the headspace of the vials for equilibration.

Equilibrium is reached at a temperature typically of  $30 \pm 1^\circ\text{C}$  after a minimum period of 18 hours.

After the equilibration process is completed the resulting  $\text{CO}_2$  is transferred by means of the continuous flow system into the ion source of the mass spectrometer.  $\text{CO}_2$  reference gas is also introduced into the IRMS by means of the continuous flow system. The measurement is carried out according to a specific protocol for each kind of equipment.

## 9. Calculation

The intensities for  $m/z = 44, 45, 46$  are recorded for each sample and reference materials analysed in a batch of measurements. The  $^{18}\text{O}/^{16}\text{O}$  isotope ratios are then calculated by the computer and the software of the IRMS instrument according to the principles explained in section 6.1. In practice the  $^{18}\text{O}/^{16}\text{O}$  isotope ratios are measured against a working standard previously calibrated against the V-SMOW. Small variations may occur while measuring on line due to changes in the instrumental conditions. In such a case the  $\delta^{18}\text{O}$  of the samples must be corrected according to the difference in the  $\delta^{18}\text{O}$  value from the working standard and its assigned value, which was calibrated beforehand against V-SMOW. Between two measurements of the working standard, the variation is the correction applied to the sample results that may be assumed to be linear. Indeed, the working standard must be measured at the beginning and at the end of all sample series. Therefore a correction can be calculated for each sample using linear interpolation between two values (the difference between the assigned value of the working standard and the measurements of the obtained values).

The final results are presented as relative  $\delta^{18}\text{O}_{\text{V-SMOW}}$  values expressed in ‰.

**COMPENDIUM OF INTERNATIONAL METHODS OF WINE AND MUST ANALYSIS**  
**Method for 18O/16O isotope ratio determination of water in wines and must (Type-II)**

---

□  $^{18}\text{O}_{V-SMOW}$  values are calculated using the following equation:

$$\delta^{18}\text{O}_{V-SMOW} = \left[ \frac{\left(\frac{^{18}\text{O}}{^{16}\text{O}}\right)_{\text{sample}} - \left(\frac{^{18}\text{O}}{^{16}\text{O}}\right)_{V-SMOW}}{\left(\frac{^{18}\text{O}}{^{16}\text{O}}\right)_{V-SMOW}} \right] \times 1000 [‰]$$

The □  $^{18}\text{O}$  value normalized versus the V-SMOW/SLAP scale is calculated using the following equation:

$$\delta^{18}\text{O}_{V-SMOW/SLAP} = \left[ \frac{\delta^{18}\text{O}_{\text{sample}} - \delta^{18}\text{O}_{V-SMOW}}{\delta^{18}\text{O}_{V-SMOW} - \delta^{18}\text{O}_{SLAP}} \right] \times 55.5 [‰]$$

The □  $^{18}\text{O}_{V-SMOW}$  value accepted for SLAP is -55.5‰ (see also 5.1).

### 10. Precision

The repeatability (r) is equal to 0.24 ‰.

The reproducibility (R) is equal to 0.50 ‰.

#### Summary of statistical results

|                  | General average (‰) | Standard deviation of repeatability (‰) $s_r$ | Repeatability (‰) r | Standard deviation of reproducibility (‰) $s_R$ | Reproducibility (‰) R |
|------------------|---------------------|---|---------------------|---|-----------------------|
| <b>Water</b>     |                     |   |                     |   |                       |
| Sample 1         | -8.20               | 0.068   | 0.19                | 0.171   | 0.48                  |
| Sample 2         | -8.22               | 0.096   | 0.27                | 0.136   | 0.38                  |
| <b>Wine N° 1</b> |                     |   |                     |   |                       |
| Sample 5         | 6.87                | 0.098   | 0.27                | 0.220   | 0.62                  |
| Sample 8         | 6.02                | 0.074   | 0.21                | 0.167   | 0.47                  |

**COMPENDIUM OF INTERNATIONAL METHODS OF WINE AND MUST ANALYSIS**  
**Method for 18O/16O isotope ratio determination of water in wines and must (Type-II)**

---

|                  |       |       |      |       |      |
|------------------|-------|-------|------|-------|------|
| Sample 9         | 5.19  | 0.094 | 0.26 | 0.194 | 0.54 |
| Sample 4         | 3.59  | 0.106 | 0.30 | 0.205 | 0.57 |
| <b>Wine N° 2</b> |       |       |      |       |      |
| Sample 3         | -1.54 | 0.065 | 0.18 | 0.165 | 0.46 |
| Sample 6         | -1.79 | 0.078 | 0.22 | 0.141 | 0.40 |
| Sample 7         | -2.04 | 0.089 | 0.25 | 0.173 | 0.49 |
| Sample 10        | -2.61 | 0.103 | 0.29 | 0.200 | 0.56 |

**11. Inter-laboratories studies**

Bulletin de l'O.I.V. janvier-février 1997, 791-792, p.53 - 65.

**12. Bibliography**

- [1] Allison, C.E., Francey, R.J. and Meijer., H.A., (1995) Recommendations for the Reporting of Stable Isotopes Measurements of carbon and oxygen. Proceedings of a consultants meeting held in Vienna, 1 - 3. Dec. 1993, *IAEA-TECDOC-825*, 155-162, Vienna, Austria.
- [2] Baertschi, P., (1976) Absolute <sup>18</sup>O Content of Standard Mean Ocean Water. *Earth and Planetary Science Letters*, 31, 341-344.
- [3] Breas, O., Reniero, F. and Serrini, G., (1994) Isotope Ratio Mass Spectrometry: Analysis of wines from different European Countries. *Rap. Comm. Mass Spectrom.*, 8, 967-987.
- [4] Craig, H., (1957) Isotopic standards for carbon and oxygen and correction factors for mass spectrometric analysis of carbon dioxide. *Geochim. Cosmochim. Acta*, 12, 133-149.
- [5] Craig, H., (1961) Isotopic Variations in Meteoric Waters. *Science*, 133,



# COMPENDIUM OF INTERNATIONAL METHODS OF WINE AND MUST ANALYSIS

## Method for $^{18}\text{O}/^{16}\text{O}$ isotope ratio determination of water in wines and must (Type-II)

---

- 1702-1703[6] Craig, H., (1961) Standard for reporting concentrations of deuterium and oxygen-18 in natural waters. *Science*, 133, 1833-1834.
- [7] Coplen, T., (1988) Normalization of oxygen and hydrogen data. *Chemical Geology* (Isotope Geoscience Section), 72, 293-297
  - [8] Coplen, T. and Hopple, J., (1995) Audit of V-SMOW distributed by the US National Institute of Standards and Technology. Proceedings of a consultants meeting held in Vienna, 1 - 3. Dec. 1993, *IAEA-TECDOC-825*, 35-38 IAEA, Vienna, Austria.
  - [9] Dunbar, J., (1982 Detection of added water and sugar in New Zealand commercial wines.). Elsevier Scientific Publishing Corp. Edts. Amsterdam, 1495-501.
  - [10] Epstein, S. and Mayeda, T. (1953) Variations of the  $^{18}\text{O}/^{16}\text{O}$  ratio in natural waters. *Geochim. Cosmochim. Acta*, 4, 213 .
  - [11] Förstel, H. (1992) Projet de description d'une méthode : variation naturelle du rapport des isotopes  $^{16}\text{O}$  et  $^{18}\text{O}$  dans l'eau comme méthode d'analyse physique du vin en vue du contrôle de l'origine et de l'addition d'eau. *OIV, FV n° 919*, 1955/220792.
  - [12] Gonfiantini, R., (1978) Standards for stable isotope measurements in natural compounds. *Nature*, 271, 534-536.
  - [13] Gonfiantini, R., (1987) Report on an advisory group meeting on stable isotope reference samples for geochemical and hydrochemical investigations. IAEA, Vienna, Austria.
  - [14] Gonfiantini, R., Stichler, W. and Rozanski, K., (1995) Standards and Intercomparison Materials distributed by the IAEA for Stable Isotopes Measurements. Proceedings of a consultants meeting held in Vienna, 1 - 3. Dec. 1993, *IAEA-TECDOC-825*, 13-29 Vienna, Austria.
  - [15] Guidelines for Collaborative Study Procedures (1989) *J. Assoc. Off. Anal. Chem.*, 72, 694-704.
  - [16] Martin, G.J., Zhang, B.L., Day, M. and Lees, M., (1993) Authentification des vins et des produits de la vigne par utilisation conjointe des analyses élémentaire et isotopique. *OIV, F.V.*, n°917, 1953/220792.
  - [17] Martin, G.J., Förstel, H. and Moussa, I. (1995) La recherche du mouillage

COMPENDIUM OF INTERNATIONAL METHODS OF WINE AND MUST ANALYSIS  
Method for  $^{18}\text{O}/^{16}\text{O}$  isotope ratio determination of water in wines and must (Type-II)

---

- des vins par analyse isotopique  $^2\text{H}$  et  $^{18}\text{O}$ . *OIV, FV n° 1006*, 2268/240595
- [18] Martin, G.J. (1996) Recherche du mouillage des vins par la mesure de la teneur en  $^{18}\text{O}$  de l'eau des vins. *OIV, FV n° 1018*, 2325/300196.
  - [19] Martin, G.J. and Lees, M., (1997) Détection de l'enrichissement des vins par concentration des moûts au moyen de l'analyse isotopique  $^2\text{H}$  et  $^{18}\text{O}$  de l'eau des vins. *OIV, FV n° 1019*, 2326/300196.
  - [20] Moussa, I., (1992) Recherche du mouillage dans les vins par spectrométrie de masse des rapports isotopiques (SMRI). *OIV, FV n°915*, 1937/130592.
  - [21] Werner, R.A. and Brand, W., (2001) Reference Strategies and techniques in stable isotope ratio analysis. *Rap. Comm. Mass Spectrom.*, 15, 501-519.
  - [22] Zhang, B.L., Fourel, F., Nault, N. and Martin, G.J., (1992) Influence de l'expérimentation et du traitement de l'échantillon sur la précision et la justesse des mesures des rapports isotopiques (D/H) et ( $^{18}\text{O}/^{16}\text{O}$ ). *OIV, F.V. n° 918*, 1954/220792.