

Determining the presence and content of polychlorophenols and polychloroanisols in wines, cork stoppers, wood and bentonites used as atmospheric traps (Type-IV)

---

## **OIV-MA-AS315-17 Determining the presence and content of polychlorophenols and polychloroanisols in wines, cork stoppers, wood and bentonites used as atmospheric traps**

Type IV method

### **1. Scope**

All wines, cork stoppers, bentonites (absorption traps) and wood.

### **2. Principle**

Determination of 2,4,6-trichloroanisole, 2,4,6-trichlorophenol, 2,3,4,6-tetrachloroanisole, 2,3,4,6-tetrachlorophenol, pentachloroanisole and pentachlorophenol by gas chromatography, by injecting a hexane extract of the wine and an ether/hexane extract of the solid samples to be analyzed and internal calibration.

### **3. Reagents**

Preliminary remark: all the reagents and solvents must be free of the compounds to be determined listed in 2 at the detection limit.

3.1. Purity of hexane > 99 %

3.2. Purity of ethylic ether > 99 %

3.3. Ether/hexane mixture (50/50; v/v)

3.4. or 2,5-dibromophenol purity  $\geq$  99 %

3.5. Pure ethanol

3.6. Pure deionized water, TCA free, type II in accordance with ISO standard EN 3696

3.7. 50 % vol. aqueous-alcoholic solution. Place 100 ml of absolute ethanol (3.<5) in a graduated 200-ml flask (4.9.9), add 200 ml of deionized water (3.6), and homogenize.

3.8. Internal standard:

3.8.1. 200 mg/l stock solution. Place 20 mg of internal standard (3.4) in a graduated 100-ml flask (4.9.8), add the 50 % volume aqueous-alcoholic solution (3.7) and homogenize.

3.8.2. Internal standard solution (2 mg/l). Place 1 ml of the stock solution of internal standard (3.8.1) in a graduated 100-ml flask (4.9.8), add the 50% vol aqueous-alcoholic solution (3.7) and homogenize.

# COMPENDIUM OF INTERNATIONAL METHODS OF WINE AND MUST ANALYSIS

## Determining the presence and content of polychlorophenols and polychloroanisols in wines, cork stoppers, wood and bentonites used as atmospheric traps (Type-IV)

---

3.8.3. Internal standard solution (20 µg/l). Place 1 ml of stock solution of internal standard (3.8.2) in a 100 ml graduated flask (4.9.8), add with 50 % vol aqueous-alcoholic solution

3.9. Pure products

- 2,4,6-trichloroanisole:  $\geq$  99 %, case: 87-40-1
- 2, 4, 6-trichlorophenol:  $\geq$  99.8 %, case: 88-06-2
- 2,3,5,6-tetrachloroanisole:  $\geq$  99 %, case: 6936-40-9 (note: the product sought in the samples is 2,3,4,6-tetrachloroanisole but it does not exist on the market)
- 2, 3, 4, 6-tetrachlorophenol:  $\geq$  99 %, case: 58-90-2
- pentachloroanisole:  $\geq$  99 %, case: 1825-21-1
- pentachlorophenol: 99 %, case: 87-86-5

10. Reagents for derivatisation - Piridine: acetic anhydride (1:0,4) vol.

1. Piridine:  $\geq$  99 %
2. Acetic anhydride:  $\geq$  98 %

11. Calibration stock solution at 200 mg/l

In a graduated 100-ml flask (4.9.8), place approximately 20 mg of the pure reference products (3.9.1 to 3.9.6) but whose exact weight is known (4.7), add absolute ethanol (3.5). Homogenize.

3.12. Intermediate calibration solution at 200 µg/l

In a graduated 100-ml flask (4.9.8) filled with absolute ethanol (3.5), add 100 µl of the calibration stock solution at 200 mg/l (3.11) using the 100- µl micro-syringe (4.9.1) and homogenize.

3.13. Calibration surrogate solution at 4 µg/l

In a graduated 50-ml flask (4.9.7) containing 50 % vol aqueous-alcoholic solution (3.7) add 1 ml of the intermediate calibration solution at 200 µg/l (3.11) using a 1-ml pipette (4.9.6). Add to volume 50 ml with pure ethanol (3.5) and homogenize.

3.14. Calibration solutions. It is possible to prepare various standard solutions with various concentrations by adding, using the 100- µl micro-syringe of (4.9.1), for example 50 µl of the surrogate calibration solution at 4 µg/l (3.12) to 50 ml of wine to enrich it with 4 ng/l of the substances to be determined.

# COMPENDIUM OF INTERNATIONAL METHODS OF WINE AND MUST ANALYSIS

## Determining the presence and content of polychlorophenols and polychloroanisols in wines, cork stoppers, wood and bentonites used as atmospheric traps (Type-IV)

---

The same reasoning can be used to prepare calibration solutions of various concentrations, either using aqueous-alcoholic solutions, or wine, or to enrich an extraction medium with a known quantity of pure products.

3.15. Commercially available Bentonite.

### 4. Apparatus

4.1. Gas phase chromatograph with Split-splitless injector coupled to an electron capture detector. (It is likewise possible to use a mass spectrometer)

4.2. Capillary tube of non-polar steady-state phenylmethylpolysiloxane type: (0.32 mm x 50 m, thickness of film 0.12 µm or the equivalent)

4.3. Chromatographic conditions, as an example:

4.3.1. Injection in "split-splitless" mode (valve closing time 30 seconds)

4.3.2. Carrier gas flow rate: 30 ml/min including 1 ml in the column Hydrogen U <sup>®2</sup>  
(It is likewise possible to use helium)

4.3.3. Auxiliary gas flow rate: 60 ml/min – Nitrogen with chromatographic purity ( $\geq 99,9990\%$ ). It is also possible to use argon methane.

4.3.4. Furnace gradient temperature for information purposes:

- from 40 °C to 160 °C at a rate of 2 °C/min
- from 160 °C to 200 °C at a rate of 5 °C/min
- step at 220 °C for 10 min

5. Injector temperature: 250 °C

6. Detector temperature: 250 °C

4. Acquisition and integration: acquisition is by computer. The peaks of the various compounds identified by comparison with the reference are then integrated.

5. Magnetic agitator.

6. Vortex with adaptation for 30-ml flask (4.9.3)

7. Precision balance to within 0.1 mg

8. Manual or electric household grate

9. Laboratory equipment:

1. 100- µl micro-syringe

# COMPENDIUM OF INTERNATIONAL METHODS OF WINE AND MUST ANALYSIS

## Determining the presence and content of polychlorophenols and polychloroanisols in wines, cork stoppers, wood and bentonites used as atmospheric traps (Type-IV)

---

2. 10- µl micro-syringe
3. 30-ml flask closing with a screwed plug and cover with one side Teflon-coated
4. 10-ml stick pipette graduated 1/10 ml
5. 5-ml stick pipette graduated 1/10 ml
6. 1-ml precision pipette
7. Graduated 50-ml flask
8. Graduated 100-ml flask
9. Graduated 200-ml flask
10. 10 100-ml separating funnel
11. Pasteur pipettes and suitable propipette pear
12. Household aluminum foil, roll-form.
13. Centrifuge

### 5. Sample preparation

- 5.1. The stopper is grated (4.8) or cut into pieces (dimension < 3 mm)
- 5.2. Wood is cut with a clipper to obtain pieces (dimension < 3 mm)
- 5.3. The bentonite (3.15) (30 g for example) is spread out over a strip of aluminum foil (4.9.12) of approximately 30 cm x 20 cm and is exposed to the atmosphere to be analyzed for at least 5 days.

### 6. Operating method

- 6.1. Extraction process for solid samples:
  - 6.1.1. Stopper: in a 30-ml flask (4.9.3), place approximately 1 g of grated stopper (5.1) but of a precisely known weight (4.7)
  - 6.1.2. Wood: in a 30-ml flask (4.9.3), place approximately 2 g of wood chips (5.2) but of a precisely known weight (4.7)
  - 6.1.3. Control Bentonite: in a 30-ml flask (4.9.3), place approximately 5 g of bentonite (3.15) but of a precisely known weight (4.7)
  - 6.1.4. Sample bentonite: in a 30-ml flask (4.9.3), place approximately 5 g of bentonite (5.3) of a precisely known weight (4.7)

# COMPENDIUM OF INTERNATIONAL METHODS OF WINE AND MUST ANALYSIS

## Determining the presence and content of polychlorophenols and polychloroanisols in wines, cork stoppers, wood and bentonites used as atmospheric traps (Type-IV)

---

- 6.1.5. Add 10 ml (4.9.4) of ether/hexane mixture (3.3)
- 6.1.6. Add with the micro-syringe (4.9.1) 50  $\mu$ l of the internal standard solution (3.8.2)
- 6.1.7. Agitate with the vortex (4.6) for 3 min
- 6.1.8. Recover the ether/hexane liquid phase in a 30-ml flask (4.9.3)
- 6.1.9. Repeat the extraction operation on the sample with 2 times 5 ml of ether/hexane mixture (3.3)
- 6.1.10. Final extract: mix the 3 phases of ether/hexane.
- 6.2. Extraction of the wine and calibration solution
  - 6.2.1. Sample 50 ml of wine or calibration solution (using the graduated flask (4.9.7)
  - 6.2.2. Place them in the 100-ml graduated flask (4.9.8)
  - 6.2.3. Add with the microsyringe (4.9.1) 50  $\mu$ l of internal standard (3.8.3)
  - 6.2.4. Add 4 ml (4.9.5) of hexane (3.1)
  - 6.2.5. Carry out the extraction using the magnetic stirrer (4.5) for 5 min.
  - 6.2.6. Elutriate into the funnel (4.9.10)
  - 6.2.7. Recover the organic phase with the emulsion in a 30-ml flask (4.9.3) and aqueous phase in the 100-ml graduated flask (4.9.8)
  - 6.2.8. Repeat the extraction of the wine or calibration solution using 2 ml of hexane (3.1)
  - 6.2.9. Carry out the extraction using the magnetic stirrer (4.5) for 5 min.
  - 6.2.10. Elutriate into the funnel (4.9.10)
  - 6.2.11. Recover the organic phase with the emulsion in the same 30-ml flask mentioned in 6.2.7 (containing the organic phase obtained upon the first extraction)
  - 6.2.12. Break the emulsion of the organic phase by centrifugation (4.9.13) by eliminating the lower aqueous phase using a Pasteur pipette (4.9.11) fitted with a propipette pear.
  - 6.2.13. Final wine extract and calibration solutions: the residual organic extract
- 6.3. Analyze:
  - 6.3.1. Add final extract (6.1.11 or 6.2.13) 100  $\mu$ l (4.9.1) of the pyridine acetic anhydride reagent mixture (3.10) for the derivatisation.
  - 6.3.2. Mix using a magnetic stirrer (4.5) for 10 min.
  - 6.3.3. Inject 2  $\mu$ l of derivatised final extract (6.3.2) into the chromatograph

## 7. Calculation

# COMPENDIUM OF INTERNATIONAL METHODS OF WINE AND MUST ANALYSIS

## Determining the presence and content of polychlorophenols and polychloroanisols in wines, cork stoppers, wood and bentonites used as atmospheric traps (Type-IV)

---

$$\text{Concentration of product} = \frac{\text{Product peak area}}{\text{Peak area of internal standard}} * \text{Response factor}$$

Response factor = concentration of calibration solution (3.13) \* (Peak area of the internal standard / \*(Peak area of the pure product in the calibration solution).

Check the calibration by ensuring the response factors +/- 10 %.

### 8. Results

The results are expressed in ng/l for the wine and ng/g for the cork stoppers, bentonites and wood.

### 9. Characteristics of the method

#### 9.1. Coverage rate

The coverage rate calculated in relation to the quantities added in terms of wood chips, polychloroanisols and polychlorophenols of 115 ng/g is:

- 2,4,6-trichloroanisol: 96 %
- 2,4,6-trichlorophenol: 96 %
- 2,3,4,6-tetrachloroanisol: 96 %
- 2,3,4,6-tetrachlorophenol: 97 %
- pentachloroanisol: 96 %
- pentachlorophenol: 97 %

#### 9.2. Measurement repeatability

Calculated for each product, the uncertainties are as follows:

In a stopper ng/g	Mean	Standard deviation	Repeatability
2,4,6-trichloroanisol	1.2	0.1	0.28
2,4,6-trichlorophenol	26	3.3	9.24
2,3,4,6-tetrachloroanisol	1.77	0.44	1.23
2,3,4,6-tetrachlorophenol	2.59	0.33	0.92

# COMPENDIUM OF INTERNATIONAL METHODS OF WINE AND MUST ANALYSIS

## Determining the presence and content of polychlorophenols and polychloroanisols in wines, cork stoppers, wood and bentonites used as atmospheric traps (Type-IV)

pentachloroanisol	23.3	2.9	8.12
pentachlorophenol	7.39	1.91	5.35
In wood with 23 ng/g	Standard deviation	Repeatability	
2,4,6-trichloroanisol	1.9	5.3	
2,4,6-trichlorophenol	1.9	5.3	
2,3,4,6-tetrachloroanisol	2.6	7.4	
2,3,4,6-tetrachlorophenol	3.3	9.3	
pentachloroanisol	2.7	7.5	
pentachlorophenol	3.6	10.1	
In wine with 10 ng/l	Standard deviation	Repeatability	
2,4,6-trichloroanisol	0,4	1,1	
2,4,6-trichlorophenol	2,1	5,9	
2,3,4,6-tetrachloroanisol	0,6	1,7	
2,3,4,6-tetrachlorophenol	4	11,2	
pentachloroanisol	1,2	3,4	
pentachlorophenol	6,5	18,2	
In bentonite with 15 ng/g	Standard deviation	Repeatability	
2,4,6-trichloroanisol	0,9	2,5	
2,4,6-trichlorophenol	4	11,2	

## COMPENDIUM OF INTERNATIONAL METHODS OF WINE AND MUST ANALYSIS

### Determining the presence and content of polychlorophenols and polychloroanisols in wines, cork stoppers, wood and bentonites used as atmospheric traps (Type-IV)

2,3,4,6-tetrachloroanisol	1,2	3,4
2,3,4,6-tetrachlorophenol	5,2	14,6
pentachloroanisol	4,3	12,0
pentachlorophenol	12,1	33,9

9.3. Detection limits (DL) and quantification limits (QL) calculated according to the OIV method:

#### 9.3.1. Wood

	DL in ng/g	QL in ng/g
2,4,6-trichloroanisol	0.72	2.4
2,4,6-trichlorophenol	0.62	2.0
2,3,4,6-tetrachloroanisol	0.59	2.0
2,3,4,6-tetrachlorophenol	1.12	3.74
pentachloroanisol	0.41	1.4
pentachlorophenol	0.91	3.1

#### 9.3.2. Bentonite

	DL in ng/g	QL in ng/g
2,4,6-trichloroanisol	0.5	1
2,4,6-trichlorophenol	1	3
2,3,4,6-tetrachloroanisol	0.5	1
2,3,4,6-tetrachlorophenol	1	3
pentachloroanisol	0.5	1



## COMPENDIUM OF INTERNATIONAL METHODS OF WINE AND MUST ANALYSIS

### Determining the presence and content of polychlorophenols and polychloroanisols in wines, cork stoppers, wood and bentonites used as atmospheric traps (Type-IV)

	pentachlorophenol	Not det.	Not det.
9.3.3.	Stopper		
		DL in ng/g	QL in ng/g
	2,4,6-trichloroanisol	0.5	1.5
	2,4,6-trichlorophenol	1	2
	2,3,4,6-tetrachloroanisol	0.5	1.5
	2,3,4,6-tetrachlorophenol	1	2
	pentachloroanisol	0.5	1.5
	pentachlorophenol	1	2
9.3.4.	Wine		
		DL in ng/l	QL in ng/l
	2,4,6-trichloroanisol	0.3	1
	2,4,6-trichlorophenol	1	3
	2,3,4,6-tetrachloroanisol	0.3	1
	2,3,4,6-tetrachlorophenol	0.3	1
	pentachloroanisol	0.5	3
	pentachlorophenol	1	3

®<sup>2</sup> Air Liquide